

**AUTOMATED
SOFTENING
POINT
TESTER
Model
ASP-5**



ASP-5 automates the softening point test by Ring-and-Ball method for bitumen and similar substances, prescribed in ASTM D36.

A photo-electric device with wide light beam is employed to assure a reliable automatic detection of softening point.

EASY OPERATION: Just set up the specimen, determine the bath liquid type, and press the START key; all the rest of the test procedures such as heating, softening point detection, and heater cooling are performed automatically.

EXACT TEMPERATURE RISE CONTROL: Exact temperature rise rate can be attained not only with water bath but also with glycerin bath by the precise heater control.

RELIABLE FALLING BALL DETECTION: Thanks to the wide light projector and wide light receiver, the ASP-5 detects the falling ball even if the ball falls with some angle.

SAFETY MECHANISM: Generation of harmful vapor from overheated glycerin is prevented by the safety shutdown.

CALIBRATION CAPABILITY: By simply setting a traceable thermometer right next to the PT-100 temperature sensor and running a test to record and compare the two temperature readings, user can easily carry out a traceable temperature calibration.

DATA HANDLING: The unit is equipped to store up to 20 successive test results and export to an optional printer or to a LIMS through an RS232 port.

COMPACT DESIGN: The test head unit can be stacked on the control unit and is also small enough (only 23cm in width) to be operated inside of a fume hood.

SPECIFICATIONS:

CONFORMING STANDARDS:

ASTM D36 etc.

TYPE:

Two-Rings & Two-Balls type

MEASURING RANGE:

Ambient to 200°C (392 °F)

TEMPERATUR SENSOR:

PT-100 in stainless steel sheath

SOFTENING POINT DETECTOR:

Light projector: Cylindrical lamp

Light receiver: Photo-diode array

DISPLAY:

Fluorescent display tube (VFD Module)

HEATER:

Nichrom coil heater: 1kW @100V or 1kW @220V

SEQUENSE CONTROL:

By an on-board microcomputer

TEST MODE:

Water bath mode and Glycerin bath mode

DATA OUTPUT:

RS-232C 1 channel (for PC or Optional Printer)

DATA MEMORY:

Last 20 sets of test results with sample ID

SAFETY MECHANISM:

Automatically shuts off and the problem is reported by buzzer and display, in case:

- (a) the sample falls on only one side and not on the other within the set temperature range,
- (b) the sample does not fall on either side within the set upper limit temperature,
- (c) light projector or receiver is found defective,
- (d) temperature sensor is found defective,
- (e) temperature sensor not placed in the right position,
- (f) built-in battery is found drained out, or
- (g) control computer runs away (no display)

DIMENSION & WEIGHT:

Control unit: 230W x 455D x 110H (mm), 6.0kg

Test unit: 230W x 380D x 270H (mm), 7.0kg

INSTALLATION SITE:

Ambient temperature: 0 to 40 °C

RH: Less than 90%(no condensation)

POWER SUPPLY:

AC100/120V or 220/240V 50/60Hz

(set at the factory, specify when ordering) 1.5kW(max.)

ORDERING INFORMATION:

STANDARD ACCESSORIES:

- | | |
|--|-------|
| 1. Ring Holder and Bottom Plate Assembly | 1pc |
| 2. Shouldered Ring | 6pcs |
| 3. Ball, 3/8in. | 6pcs |
| 4. Ball Centering Guide | 2pcs |
| 5. Beaker | 1 pc |
| 6. Beaker Handle | 1pc |
| 7. Temperature Sensor Assy | 1pc |
| 8. Temperature Sensor Holder | 1pc |
| 9. Power Connecting Cable, 0.6m | 1pc |
| 10. Signal Connecting Cable, 0.6m | 1pc |
| 11. AC Power Cord, 3.0m (<AC125V)
or 2.5m (>AC200V) | 1 pc |
| 12. Instruction Manual | 1copy |

OPIONAL ACCESSORIES:

Power Connecting Cable, 3m

Signal Connecting Cable, 3m

Built-in Clock Board (for time/date verification)

Printer, BS2-80TS (we recommend that optional Built-in Clock Board be purchased with this Printer)

SUGGESTED SPARES:

- | | |
|--|-------|
| 1. Ring Holder and Bottom Plate Assembly | 1pc |
| 2. Nichrom Heater Coil 1kW @100V
or 1kW @220V | 1pc |
| 3. Light Projector Lamp H-0937 | 1pc |
| 4. Light Receiver Assy | 1pc |
| 5. Temperature Sensor Assy | 1pc |
| 6. Shouldered Ring | 20pcs |
| 7. Ball, 3/8in. | 50pcs |
| 8. Ball Centering Guide | 2pc |
| 9. Beaker | 1 pc |
| 10. Beaker Handle | 1pc |

Specifications subject to change without prior notice.

TANAKA SCIENTIFIC LIMITED

7-10-3, Ayase, Adachi-ku, Tokyo 120-0005 Japan

Tel: +81-3-3620-1711 Fax: +81-3-3620-1713

URL: <http://www.tanaka-sci.com>

e-mail: sales@tanaka-sci.com Printed in Japan 0909(E)